

# Chlorine Test Kit

1 drop = 5 ppm / 10 mL

# TK4010-Z

orange caps

## KIT COMPONENTS:

SA1940-A	Sulfuric Acid 50%, 30 mL
PI1450-B	Potassium Iodide 50%, 60 mL
ST5010-B	Starch Indicator Solution 1%, 60 mL
ST2705-B	Sodium Thiosulfate 0.0365N, 60 mL
SY-2012-P	Syringe, 12 mL
VL-0525-V	Vial, 5-25 mL

**INTERFERENCES:** All oxidizers, including Bromine, are positive interferences. Interferences include a pH over 8, Total Hardness over 1000 ppm, Sulfate over 1000 ppm, Total Alkalinity over 150 ppm, any concentration of Nitrite, Nitrate over 200 ppm, Silica Dioxide over 50 ppm, Copper over 10 ppm, any concentration of Ferrous Iron (Fe<sup>2+</sup>) and Ferric Iron (Fe<sup>3+</sup>) over 5 ppm.

## SAFETY TIPS:



Wear  
Gloves



Use Eye  
Protection



Read  
SDS

## TESTING TIPS:



Collect  
Accurate  
Sample



Hold  
Bottles  
Vertically



Ensure  
Proper  
Lighting

**ATTENTION:** As necessary, calibrate this kit against a known standard made with plant / make-up water. Be sure to collect a representative sample.

Video Procedure




**1** Rinse vial three times with sample to be tested. **Fill vial to 10 mL.**


*Note: It is important that each reagent be added and then mixed well for at least 5 seconds before the addition of the subsequent reagent.*

**2** Add 10 drops of **Potassium Iodide** (PI1450) and swirl 5 seconds to mix.

**3** Add 5 drops of **Starch Indicator Solution** (ST5010) and swirl 5 seconds to mix.



STEP 1



STEP 3

**4** Add 3 drops of **Sulfuric Acid 50%** (SA1940) and swirl 5 seconds to mix. The sample will turn a blue-black color.


(In caustic solutions, more drops may be necessary.)



STEP 4

**5** Add **Sodium Thiosulfate 0.0365N** (ST2705) one drop at a time while swirling. Count the number of drops until the sample returns to its original color.

# drops x 5 = ppm Chlorine



STEP 5

# Chlorine Test Kit

1 drop = 10 ppm / 5 mL

# TK4010-Z

orange caps

## KIT COMPONENTS:

SA1940-A	Sulfuric Acid 50%, 30 mL
PI1450-B	Potassium Iodide 50%, 60 mL
ST5010-B	Starch Indicator Solution 1%, 60 mL
ST2705-B	Sodium Thiosulfate 0.0365N, 60 mL
SY-2012-P	Syringe, 12 mL
VL-0525-V	Vial, 5-25 mL

**INTERFERENCES:** All oxidizers, including Bromine, are positive interferences. Interferences include a pH over 8, Total Hardness over 1000 ppm, Sulfate over 1000 ppm, Total Alkalinity over 150 ppm, any concentration of Nitrite, Nitrate over 200 ppm, Silica Dioxide over 50 ppm, Copper over 10 ppm, any concentration of Ferrous Iron (Fe<sup>2+</sup>) and Ferric Iron (Fe<sup>3+</sup>) over 5 ppm.

## SAFETY TIPS:



Wear  
Gloves



Use Eye  
Protection



Read  
SDS

## TESTING TIPS:



Collect  
Accurate  
Sample



Hold  
Bottles  
Vertically



Ensure  
Proper  
Lighting

**ATTENTION:** As necessary, calibrate this kit against a known standard made with plant / make-up water. Be sure to collect a representative sample.

Video Procedure




**1** Rinse vial three times with sample to be tested. **Use the syringe to add 5 mL of the sample to the vial. Then, fill vial to 10 mL with tap water.**


*Note: It is important that each reagent be added and then mixed well for at least 5 seconds before the addition of the subsequent reagent.*

**2** Add 10 drops of **Potassium Iodide** (PI1450) and swirl 5 seconds to mix.

**3** Add 5 drops of **Starch Indicator Solution** (ST5010) and swirl 5 seconds to mix.



STEP 1



STEP 3

**4** Add 3 drops of **Sulfuric Acid 50%** (SA1940) and swirl 5 seconds to mix. The sample will turn a blue-black color.


(In caustic solutions, more drops may be necessary.)



STEP 4

**5** Add **Sodium Thiosulfate 0.0365N** (ST2705) one drop at a time while swirling. Count the number of drops until the sample returns to its original color.

# drops x 10 = ppm Chlorine



STEP 5