

Cresol Red pH Test Kit

pH 7.2 – 8.6 (0.2 increments)

TK9906-Z
white caps

KIT COMPONENTS:

CR2702-B	Cresol Rd 0.02%, 60 mL
0230	Test Tube, 5 mL, w/cap
CP-0020-DR	Dropper, 0.5-1.0 mL
2186	Cresol Red pH Comparator

SAFETY TIPS:



Wear
Gloves



Use Eye
Protection



Read
SDS

TESTING TIPS:



Collect
Accurate
Sample



Hold
Bottles
Vertically




Ensure
Proper
Lighting

INTERFERENCES: Turbid or highly colored samples may mask the color change at the endpoint.

ATTENTION: As necessary, calibrate this kit against a known standard made with plant / make-up water. Be sure to collect a representative sample.

1 Rinse test tube three times with sample to be tested. **Fill test tube to the line.** Filter sample if needed.



2 Use the 1 mL pipet to **add 0.5 mL of Cresol Red Indicator** (CR2702). Cap and mix. Wipe the test tube dry and place in the comparator.



3 Match the color of the test tube to the color standard. Record as pH units.

